



LOAD CELL CALIBRATION CERTIFICATION

CONDITION : FINAL
 MODEL: SMTM-200N-1184 SERIAL: 1620946 BRIDGE: A CAPACITY: 200 N
 PROCEDURE: C-1260
 INPUT RESISTANCE: 350.5 OHM OUTPUT RESISTANCE: 350.4 OHM
 ZERO BALANCE: -0.239 %RO

TEST CONDITIONS

TEMPERATURE: 74 °F (23 °C) HUMIDITY: 30% EXCITATION: 5 VDC

TRACEABILITY

INDICATOR BRD294
 NIST #, WEIGHTS 240815002 DUE 27-MAY-2030
 NIST #, INDICATOR EVL958572

SHUNT CALIBRATION

	Shunt (+/- .01%)	Output	Straight Line Conversion	Connections*
TENSION	60.000 KOhm	1.45615 mV/V	158.28 N	-Out to -Exc

*For models wired with +Sense, - Sense or -SCal leads, resistor connections are actually to these leads in place of +Exc, -Exc or -Out respectively.

PERFORMANCE

	Rated Output	SEB Output	Nonlinearity	Hysteresis	SEB
TENSION	1.84018 mV/V	1.83994 mV/V	-0.018 %FS	0.004 %FS	±0.013 %FS

STATIC ERROR BAND (SEB) The band of maximum deviations of the ascending and descending calibration points from a best fit straight line through zero OUTPUT. It includes the effects of NONLINEARITY, HYSTERESIS and nonreturn to MINIMUM LOAD.

TEST LOAD		RECORDED READINGS (mV/V)
APPLIED (lbf)	(N)	Tension
0	0.000	.00000
20	88.964	.81821
45	200.170	1.84174
20	88.964	.81828
0	0.000	- .00012

Interface Inc. certifies that all calibration measurements are traceable to NIST. Estimated uncertainty is 0.040% RDG. Results relate to load cell serial 1620946 only.
 DO NOT REPRODUCE THIS REPORT except in full or with Interface Inc. written approval.

TECHNICIAN : FLORES, NOEL

CALIBRATION DATE : 09-MAR-2026

APPROVED: _____